

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001414  
12:55 31/03/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:44  
PRESSURIZING 2:50  
STERILIZING 3:52  
136.0 C 220kPa 3:52  
136.4 C 226kPa 4:22  
136.4 C 225kPa 4:52  
136.4 C 226kPa 5:22  
136.4 C 222kPa 5:52  
136.4 C 225kPa 6:22  
136.2 C 224kPa 6:52  
136.2 C 224kPa 7:22  
136.2 C 224kPa 7:22  
VENTING 7:23  
AIR DRYING 8:12  
CYCLE COMPLETE 23:12  
-----

Digital Signature #  
F3F3B48EAE0AE412  
-----