

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001220  
12:35 18/08/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:58  
PRESSURIZING 3:59  
STERILIZING 5:11  
136.0 C 220kPa 5:11  
136.2 C 222kPa 5:41  
136.4 C 225kPa 6:11  
136.2 C 222kPa 6:41  
136.3 C 224kPa 7:11  
136.1 C 220kPa 7:41  
136.3 C 223kPa 8:11  
136.2 C 223kPa 8:41  
136.2 C 223kPa 8:41  
VENTING 8:41  
AIR DRYING 9:36  
CYCLE COMPLETE 24:36  
-----

Digital Signature #  
B04F027DCF0B1743  
-----