
Midmark M11 - v1.0.0
Total Cycles: 1893

Name: RITTER 1
Sterilizer ID: V2178198
05/02/2022 11:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 209.8 F 0.0
2:00 207.0 F 3.7
4:00 218.9 F 9.4
6:00 226.7 F 6.5
8:00 236.2 F 11.1
10:00 235.6 F 8.5
12:00 240.1 F 10.0
14:00 247.8 F 14.2
16:00 260.4 F 21.3
18:00 271.8 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.6 F 29.0
2:00 271.6 F 28.7
3:00 272.6 F 29.0
4:00 271.6 F 28.5

Min 271.5 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.6 F 0.0
5:00 220.8 F 0.0
10:00 238.6 F 0.0
15:00 235.7 F 0.0
20:00 223.3 F 0.0
25:00 220.7 F 0.0
30:00 221.2 F 0.0
Drying Complete

FILLING: 3:17
HEATING: 18:32
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:59
CYCLE COMPLETE

05/02/2022 12:31 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____