
Midmark M11 - v1.0.0
Total Cycles: 2611

Name: RITTER 1
Sterilizer ID: V2178198
11/28/2022 11:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.7 F 0.0
2:00 195.5 F 4.9
4:00 202.3 F 10.4
6:00 222.6 F 5.4
8:00 232.9 F 11.8
10:00 234.8 F 9.5
12:00 236.9 F 8.7
14:00 249.1 F 15.4
16:00 262.0 F 22.7
18:00 272.4 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.4 F 29.5
2:00 271.3 F 28.8
3:00 272.4 F 29.7
4:00 271.9 F 29.1

Min 271.3 F 28.8
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.3 F 0.0
5:00 207.2 F 0.0
10:00 214.4 F 0.0
15:00 217.4 F 0.0
20:00 219.2 F 0.0
25:00 220.5 F 0.0
30:00 222.3 F 0.0
Drying Complete

FILLING: 2:20
HEATING: 18:14
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:47
CYCLE COMPLETE

11/28/2022 12:16 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____