
Midmark M11 - v1.0.0
Total Cycles: 4831

Name: RITTER 1
Sterilizer ID: V2178198
08/19/2024 01:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 100.6 F 0.0
2:00 107.9 F 0.1
4:00 134.1 F 0.9
6:00 161.3 F 1.5
8:00 181.5 F 4.4
10:00 212.0 F 7.6
12:00 213.2 F 10.9
14:00 222.2 F 5.7
16:00 229.3 F 7.8
18:00 239.9 F 12.0
20:00 235.5 F 8.7
22:00 243.6 F 11.9
24:00 244.8 F 12.5
26:00 256.1 F 18.4
28:00 265.9 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.4 F 28.9
2:00 272.5 F 28.8
3:00 271.8 F 28.3
4:00 271.7 F 28.5

Min 271.4 F 28.1
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.2 F 0.0
5:00 207.1 F 0.0
10:00 203.9 F 0.0
15:00 205.5 F 0.0
20:00 185.3 F 0.0
25:00 180.8 F 0.0
30:00 179.9 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 29:56
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:08:32
CYCLE COMPLETE

08/19/2024 02:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____