
Midmark M11 - v1.0.5
Total Cycles: 4467

Name: RITTER2
Sterilizer ID: V2313228
03/27/2025 03:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.0 F 0.0
2:00 130.2 F 1.5
4:00 135.7 F 4.3
6:00 154.5 F 8.0
8:00 178.2 F 12.2
10:00 207.6 F 5.2
12:00 223.3 F 9.3
14:00 228.7 F 11.8
16:00 235.3 F 10.9
18:00 235.0 F 9.0
20:00 248.1 F 15.3
22:00 258.8 F 21.5
24:00 268.6 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 30.3
1:00 271.9 F 30.3
2:00 271.8 F 30.3
3:00 271.8 F 30.3
4:00 271.9 F 30.5

Min 271.4 F 30.2
Max 273.8 F 31.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.3 F 0.0
5:00 187.8 F 0.0
10:00 196.2 F 0.0
15:00 198.3 F 0.0
20:00 201.1 F 0.0
25:00 200.9 F 0.0
30:00 201.8 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 25:21
STERILIZING: 4:00
VENTING: 3:58
DRYING: 30:00
TOTAL CYCLE: 01:04:24
CYCLE COMPLETE

03/27/2025 05:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____