

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002291  
10:31 19/11/2024  
AIR DRYING ONLY  
60 min. MAXIMUM  
CYCLE START 0:00  
AIR DRYING 0:02  
CYCLE COMPLETE 5:05  
-----

Digital Signature #  
F8A041465A095066  
-----