

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001163  
7:33 11/02/2021

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:25  
PRESSURIZING 2:30  
STERILIZING 3:35  
136.0 C 220kPa 3:35  
136.0 C 220kPa 4:05  
136.2 C 221kPa 4:35  
136.4 C 219kPa 5:05  
136.2 C 221kPa 5:35  
136.3 C 221kPa 6:05  
136.4 C 222kPa 6:35  
136.1 C 221kPa 7:05  
136.1 C 221kPa 7:05  
VENTING 7:05  
AIR DRYING 7:57  
CYCLE COMPLETE 16:50  
-----

Digital Signature #  
47AC959779A6B27A  
-----