

-----  
Midmark M11 - v1.0.5  
Total Cycles: 640

Name: RITTER2  
Sterilizer ID: V2313228  
12/16/2021 07:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.4 F 0.0  
2:00 85.4 F 0.0  
4:00 130.0 F 0.1  
6:00 165.9 F 0.3  
8:00 169.7 F 3.5  
10:00 180.9 F 7.7  
12:00 200.6 F 12.2  
14:00 216.4 F 5.4  
16:00 230.2 F 9.5  
18:00 235.5 F 10.7  
20:00 238.9 F 11.1  
22:00 236.1 F 9.1  
24:00 249.8 F 15.5  
26:00 261.4 F 22.2  
28:00 272.0 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.9  
1:00 272.6 F 29.6  
2:00 271.7 F 28.7  
3:00 272.4 F 29.8  
4:00 272.0 F 29.2

Min 271.3 F 28.7  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.5 F 0.0  
5:00 229.2 F 0.0  
10:00 163.2 F 0.0  
15:00 157.1 F 0.0  
20:00 152.4 F 0.0  
25:00 147.8 F 0.0  
30:00 144.2 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 28:36  
STERILIZING: 4:00  
VENTING: 2:51  
DRYING: 30:00  
TOTAL CYCLE: 01:06:19  
CYCLE COMPLETE

12/16/2021 09:05 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02