
Midmark M11 - v1.0.5
Total Cycles: 3141

Name: RITTER2
Sterilizer ID: V2313228
11/17/2023 05:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	97.3 F	0.0
2:00	101.1 F	0.2
4:00	132.6 F	1.9
6:00	165.6 F	5.2
8:00	182.6 F	9.1
10:00	200.3 F	12.8
12:00	215.4 F	5.5
14:00	226.8 F	9.6
16:00	233.3 F	9.8
18:00	240.2 F	11.7
20:00	239.0 F	10.4
22:00	252.3 F	17.0
24:00	264.2 F	23.9

STERILIZING

mm:ss	Degrees	PSI
0:00	271.4 F	28.8
1:00	272.9 F	29.5
2:00	271.5 F	28.7
3:00	272.8 F	29.7
4:00	272.0 F	29.0

Min 271.4 F 28.6
Max 272.9 F 29.7

VENTING CHAMBER

mm:ss	Drying Start	Degrees	PSI
0:00		167.5 F	0.0
5:00		199.3 F	0.0
10:00		206.7 F	0.0
15:00		206.9 F	0.0
20:00		205.9 F	0.0
25:00		206.3 F	0.0
30:00		204.4 F	0.0

Drying Complete

FILLING: 1:18
HEATING: 25:60
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:04:41
CYCLE COMPLETE

11/17/2023 07:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____