
Midmark M11 - v1.0.0
Total Cycles: 2243

Name: RITTER 1
Sterilizer ID: V2178198
08/24/2022 05:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.3 F 0.0
2:00 183.9 F 0.6
4:00 210.1 F 3.7
6:00 226.2 F 9.2
8:00 235.9 F 9.5
10:00 241.5 F 12.1
12:00 240.3 F 10.8
14:00 239.2 F 9.9
16:00 254.6 F 17.6
18:00 266.9 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.4 F 28.6
2:00 272.1 F 28.7
3:00 272.0 F 28.4
4:00 272.4 F 28.9

Min 271.5 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.2 F 0.0
5:00 235.6 F 0.0
10:00 231.5 F 0.0
15:00 223.8 F 0.0
20:00 215.9 F 0.0
25:00 214.3 F 0.0
30:00 214.0 F 0.0
Drying Complete

FILLING: 1:29
HEATING: 19:24
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:02
CYCLE COMPLETE

08/24/2022 06:20 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____