
Midmark M11 - v1.0.5
Total Cycles: 1587

Name: RITTER2
Sterilizer ID: V2313228
10/12/2022 02:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.4 F 0.0
2:00 187.0 F 3.9
4:00 196.2 F 8.6
6:00 219.5 F 10.0
8:00 228.9 F 9.7
10:00 233.0 F 8.8
12:00 243.5 F 13.5
14:00 243.6 F 12.7
16:00 256.6 F 19.6
18:00 268.4 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.5 F 30.0
2:00 272.4 F 29.6
3:00 271.7 F 29.1
4:00 271.4 F 29.3

Min 271.2 F 28.9
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.9 F 0.0
5:00 232.5 F 0.0
10:00 238.0 F 0.0
15:00 231.4 F 0.0
20:00 222.2 F 0.0
25:00 221.1 F 0.0
30:00 217.6 F 0.0
Drying Complete

FILLING: 2:10
HEATING: 19:19
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:39
CYCLE COMPLETE

10/12/2022 03:45 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____