
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000634
10:31 15/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:59
PRESSURIZING 3:04
STERILIZING 3:56
136.0 C 221kPa 3:56
136.4 C 225kPa 4:26
136.4 C 221kPa 4:56
136.4 C 225kPa 5:26
136.1 C 221kPa 5:56
136.4 C 225kPa 6:26
136.3 C 222kPa 6:56
136.1 C 222kPa 7:26
136.1 C 222kPa 7:26
VENTING 7:26
AIR DRYING 8:17
CYCLE COMPLETE 8:21

Digital Signature #
A8F621D026F052BB
