
Midmark M11 - v1.0.5
Total Cycles: 4082

Name: RITTER2
Sterilizer ID: V2313228
11/04/2024 04:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.0 F 0.0
2:00 141.7 F 0.8
4:00 149.0 F 3.8
6:00 173.4 F 7.5
8:00 196.0 F 11.7
10:00 213.0 F 5.1
12:00 220.8 F 9.7
14:00 232.7 F 9.3
16:00 238.8 F 11.8
18:00 238.7 F 10.3
20:00 251.2 F 16.4
22:00 261.9 F 22.7
24:00 272.1 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.7
1:00 272.0 F 29.6
2:00 272.0 F 29.6
3:00 272.2 F 29.6
4:00 272.6 F 29.7

Min 271.9 F 29.6
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.8 F 0.0
5:00 217.6 F 0.0
10:00 210.9 F 0.0
15:00 190.8 F 0.0
20:00 188.6 F 0.0
25:00 186.8 F 0.0
30:00 187.4 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 25:13
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:04:06
CYCLE COMPLETE

11/04/2024 05:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____