
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001597
17:59 15/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:25
PRESSURIZING 2:28
STERILIZING 3:21
136.0 C 220kPa 3:21
136.3 C 224kPa 3:51
136.2 C 224kPa 4:21
136.4 C 226kPa 4:51
136.4 C 226kPa 5:21
136.4 C 226kPa 5:51
136.4 C 225kPa 6:21
136.4 C 226kPa 6:51
136.4 C 226kPa 6:51
VENTING 6:52
AIR DRYING 7:45
CYCLE COMPLETE 8:03

Digital Signature #
5A3C19728F231765
