
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001537
17:11 23/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:51
PRESSURIZING 2:52
STERILIZING 3:56
136.5 C 226kPa 3:56
136.3 C 226kPa 4:26
136.1 C 223kPa 4:56
136.4 C 225kPa 5:26
136.4 C 225kPa 5:56
136.4 C 226kPa 6:26
136.4 C 227kPa 6:56
136.1 C 223kPa 7:26
136.1 C 223kPa 7:26
VENTING 7:26
AIR DRYING 8:22
CYCLE COMPLETE 8:32

Digital Signature #
8924C296F49C9119
