

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1377

Name: RITTER 1  
Sterilizer ID: V2178198  
11/17/2021 12:58 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.7 F 0.0  
2:00 187.3 F 4.1  
4:00 197.7 F 8.5  
6:00 215.9 F 11.3  
8:00 227.9 F 9.0  
10:00 235.8 F 10.0  
12:00 240.8 F 11.9  
14:00 239.8 F 10.4  
16:00 252.9 F 17.1  
18:00 264.8 F 24.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.6  
1:00 272.5 F 29.4  
2:00 272.0 F 28.9  
3:00 271.2 F 28.7  
4:00 272.3 F 29.5  
  
Min 271.2 F 28.5  
Max 272.5 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 215.3 F 0.0  
5:00 222.5 F 0.0  
10:00 240.3 F 0.0  
15:00 244.1 F 0.0  
20:00 227.0 F 0.0  
25:00 221.5 F 0.0  
30:00 217.9 F 0.0  
Drying Complete

FILLING: 1:26  
HEATING: 19:56  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:58:38  
CYCLE COMPLETE

11/17/2021 01:56 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_