
Midmark M11 - v1.0.0
Total Cycles: 2986

Name: RITTER 1
Sterilizer ID: V2178198
03/09/2023 09:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.1 F 0.0
2:00 190.1 F 1.5
4:00 187.6 F 5.8
6:00 207.9 F 10.9
8:00 219.8 F 5.8
10:00 234.0 F 11.1
12:00 234.0 F 8.0
14:00 241.9 F 11.2
16:00 245.8 F 13.0
18:00 258.5 F 19.8
20:00 269.7 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.7 F 29.1
2:00 272.1 F 28.5
3:00 271.4 F 28.4
4:00 272.7 F 29.2

Min 271.3 F 28.1
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.0 F 0.0
5:00 224.7 F 0.0
10:00 241.9 F 0.0
15:00 238.6 F 0.0
20:00 230.3 F 0.0
25:00 224.0 F 0.0
30:00 223.2 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 21:02
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:26
CYCLE COMPLETE

03/09/2023 10:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____