
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007700
7:40 07/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:08
PRESSURIZING 3:11
STERILIZING 4:20
136.4 C 222kPa 4:20
136.4 C 222kPa 4:50
136.1 C 221kPa 5:20
136.0 C 220kPa 5:50
136.5 C 220kPa 6:20
136.3 C 222kPa 6:50
136.2 C 220kPa 7:20
136.0 C 221kPa 7:50
136.0 C 221kPa 7:50
VENTING 7:50
AIR DRYING 8:39
CYCLE COMPLETE 38:39

Digital Signature #
B4FBD74409955E31
