

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000208  
9:57 08/07/2020

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:05  
PRESSURIZING 3:09  
STERILIZING 4:08  
136.4 C 222kPa 4:08  
136.4 C 225kPa 4:38  
136.4 C 223kPa 5:08  
136.4 C 223kPa 5:38  
136.4 C 222kPa 6:08  
136.4 C 223kPa 6:38  
136.2 C 221kPa 7:08  
136.4 C 224kPa 7:38  
136.4 C 224kPa 7:38  
VENTING 7:38  
AIR DRYING 8:28  
CYCLE COMPLETE 8:37

-----

Digital Signature #  
4511C2C676553C3F  
-----