

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002443  
9:56 14/10/2021

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:55  
PRESSURIZING 1:59  
STERILIZING 2:59  
136.4 C 220kPa 2:59  
136.3 C 223kPa 3:29  
136.4 C 224kPa 3:59  
136.2 C 221kPa 4:29  
136.4 C 223kPa 4:59  
136.4 C 225kPa 5:29  
136.2 C 221kPa 5:59  
136.2 C 221kPa 6:29  
136.2 C 221kPa 6:29  
VENTING 6:29  
AIR DRYING 7:20  
CYCLE COMPLETE 7:34

-----  
Digital Signature #  
94FC7B77EE1946BB  
-----