
Midmark M11 - v1.0.5
Total Cycles: 2068

Name: RITTER2
Sterilizer ID: V2313228
02/21/2023 08:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.6 F 0.0
2:00 78.8 F 0.0
4:00 95.8 F 0.0
6:00 125.1 F 0.0
8:00 162.9 F 1.1
10:00 185.9 F 3.4
12:00 200.3 F 6.6
14:00 218.0 F 10.3
16:00 226.7 F 8.3
18:00 233.1 F 9.5
20:00 238.1 F 10.5
22:00 242.3 F 12.2
24:00 240.7 F 11.0
26:00 253.5 F 17.5
28:00 264.6 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.4 F 29.2
2:00 272.1 F 29.5
3:00 271.8 F 28.8
4:00 272.6 F 29.6

Min 271.2 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.4 F 0.0
5:00 214.1 F 0.0
10:00 191.0 F 0.0
15:00 189.5 F 0.0
20:00 189.3 F 0.0
25:00 189.3 F 0.0
30:00 202.1 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 29:53
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:07:52
CYCLE COMPLETE

02/21/2023 09:26 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____