

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3612

Name: RITTER 1  
Sterilizer ID: V2178198  
08/03/2023 12:39 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 205.5 F 0.0  
2:00 194.1 F 3.7  
4:00 208.7 F 8.8  
6:00 223.9 F 7.2  
8:00 232.8 F 9.8  
10:00 238.3 F 8.4  
12:00 244.2 F 13.3  
14:00 243.6 F 12.1  
16:00 256.5 F 19.0  
18:00 268.0 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.6 F 29.3  
2:00 271.5 F 28.4  
3:00 272.6 F 29.4  
4:00 272.1 F 28.9  
  
Min 271.3 F 28.4  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.9 F 0.0  
5:00 224.3 F 0.0  
10:00 236.4 F 0.0  
15:00 233.3 F 0.0  
20:00 229.2 F 0.0  
25:00 228.3 F 0.0  
30:00 227.1 F 0.0  
  
Drying Complete

FILLING: 1:48  
HEATING: 19:19  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:58:24  
CYCLE COMPLETE

08/03/2023 01:37 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_