

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3492

Name: RITTER 1  
Sterilizer ID: V2178198  
07/06/2023 07:37 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.5 F 0.0  
2:00 78.9 F 0.0  
4:00 84.1 F 0.0  
6:00 106.0 F 0.0  
8:00 143.8 F 0.0  
10:00 177.3 F 0.2  
12:00 201.4 F 1.0  
14:00 214.2 F 3.9  
16:00 221.7 F 6.3  
18:00 229.9 F 9.2  
20:00 238.5 F 12.8  
22:00 235.8 F 9.0  
24:00 243.0 F 11.6  
26:00 243.9 F 12.1  
28:00 241.0 F 10.6  
30:00 253.8 F 17.0  
32:00 264.5 F 23.3  
34:00 271.5 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.0  
1:00 272.7 F 28.7  
2:00 271.5 F 27.9  
3:00 272.7 F 28.8  
4:00 271.9 F 28.1  
  
Min 271.5 F 27.9  
Max 272.8 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.9 F 0.0  
5:00 218.6 F 0.0  
10:00 210.4 F 0.0  
15:00 212.1 F 0.0  
20:00 226.3 F 0.0  
25:00 239.0 F 0.0  
30:00 229.9 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 34:02  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:12:03  
CYCLE COMPLETE

07/06/2023 08:49 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_