
Midmark M11 - v1.0.0
Total Cycles: 2378

Name: RITTER 1
Sterilizer ID: V2178198
09/27/2022 12:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.2 F 0.0
2:00 163.0 F 2.9
4:00 179.5 F 7.2
6:00 199.4 F 12.4
8:00 226.4 F 7.4
10:00 235.9 F 13.0
12:00 236.4 F 10.0
14:00 233.8 F 8.1
16:00 248.0 F 14.6
18:00 260.2 F 21.4
20:00 271.2 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.3 F 29.8
2:00 271.9 F 29.1
3:00 271.3 F 29.2
4:00 273.0 F 30.3

Min 271.0 F 28.8
Max 273.7 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.4 F 0.0
5:00 210.9 F 0.0
10:00 219.4 F 0.0
15:00 226.0 F 0.0
20:00 229.5 F 0.0
25:00 229.6 F 0.0
30:00 222.6 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 20:42
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:41
CYCLE COMPLETE

09/27/2022 01:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____