
Midmark M11 - v1.0.0
Total Cycles: 4541

Name: RITTER 1
Sterilizer ID: V2178198
05/10/2024 09:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.4 F 0.0
2:00 166.3 F 2.3
4:00 176.9 F 5.5
6:00 195.1 F 9.8
8:00 227.3 F 7.6
10:00 223.7 F 7.9
12:00 235.1 F 12.5
14:00 236.1 F 9.3
16:00 239.5 F 10.1
18:00 246.3 F 13.3
20:00 257.6 F 19.3
22:00 267.8 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 272.0 F 28.5
2:00 272.2 F 28.9
3:00 271.1 F 28.3
4:00 272.1 F 28.6

Min 271.1 F 28.2
Max 272.4 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.5 F 0.0
5:00 233.1 F 0.0
10:00 213.4 F 0.0
15:00 214.5 F 0.0
20:00 211.7 F 0.0
25:00 214.9 F 0.0
30:00 209.2 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 23:21
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:02:31
CYCLE COMPLETE

05/10/2024 10:31 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____