

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3088

Name: RITTER2  
Sterilizer ID: V2313228  
11/06/2023 01:25 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 135.6 F 0.0  
2:00 137.2 F 1.4  
4:00 160.8 F 4.9  
6:00 187.3 F 9.4  
8:00 211.0 F 7.8  
10:00 225.3 F 7.6  
12:00 237.2 F 12.6  
14:00 236.6 F 9.8  
16:00 237.4 F 9.0  
18:00 249.4 F 15.2  
20:00 261.6 F 22.2  
22:00 272.8 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.8  
1:00 272.8 F 29.4  
2:00 271.4 F 28.6  
3:00 272.8 F 29.6  
4:00 272.0 F 28.9

Min 271.2 F 28.6  
Max 272.8 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.6 F 0.0  
5:00 222.8 F 0.0  
10:00 201.1 F 0.0  
15:00 199.3 F 0.0  
20:00 196.1 F 0.0  
25:00 193.3 F 0.0  
30:00 190.8 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 22:23  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:00:42  
CYCLE COMPLETE

11/06/2023 02:25 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_