
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000119
15:55 03/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:11
PRESSURIZING 4:11
STERILIZING 5:25
136.0 C 220kPa 5:25
136.4 C 225kPa 5:55
136.4 C 226kPa 6:25
136.4 C 227kPa 6:55
136.3 C 225kPa 7:25
136.4 C 226kPa 7:55
136.4 C 225kPa 8:25
136.3 C 226kPa 8:55
136.3 C 226kPa 8:55
VENTING 8:55
AIR DRYING 9:58
CYCLE COMPLETE 24:58

Digital Signature #
82287A7A3DFAAC32
