
Midmark M11 - v1.0.0
Total Cycles: 1375

Name: RITTER 1
Sterilizer ID: V2178198
11/17/2021 10:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.2 F 0.0
2:00 185.5 F 2.4
4:00 184.2 F 6.3
6:00 203.8 F 11.1
8:00 216.4 F 5.4
10:00 231.0 F 10.9
12:00 232.3 F 7.5
14:00 244.3 F 12.8
16:00 243.7 F 12.2
18:00 256.5 F 18.8
20:00 267.8 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 271.8 F 29.0
2:00 272.2 F 28.9
3:00 271.3 F 28.4
4:00 272.3 F 29.3

Min 271.3 F 28.3
Max 272.4 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 220.0 F 0.0
10:00 231.0 F 0.0
15:00 229.9 F 0.0
20:00 226.9 F 0.0
25:00 227.7 F 0.0
30:00 200.7 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 21:14
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:26
CYCLE COMPLETE

11/17/2021 11:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____