
Midmark M11 - v1.0.0
Total Cycles: 4008

Name: RITTER 1
Sterilizer ID: V2178198
11/10/2023 02:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 115.8 F 0.0
2:00 129.0 F 1.1
4:00 125.9 F 3.6
6:00 154.9 F 6.9
8:00 176.9 F 11.0
10:00 207.4 F 3.5
12:00 216.6 F 7.7
14:00 228.4 F 12.1
16:00 230.9 F 8.2
18:00 241.5 F 13.3
20:00 240.9 F 11.2
22:00 252.7 F 17.0
24:00 263.0 F 23.2
26:00 272.5 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.6 F 29.4
2:00 272.0 F 28.8
3:00 271.4 F 28.8
4:00 272.5 F 29.4

Min 271.2 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.7 F 0.0
5:00 207.9 F 0.0
10:00 215.4 F 0.0
15:00 217.3 F 0.0
20:00 217.3 F 0.0
25:00 220.5 F 0.0
30:00 219.9 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 26:19
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:04:57
CYCLE COMPLETE

11/10/2023 03:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____