

-----  
Midmark M11 - v1.0.0  
Total Cycles: 708

Name: MIDMARK  
Sterilizer ID: V2178198  
02/26/2021 11:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.0 F 0.0  
2:00 167.6 F 2.1  
4:00 188.3 F 6.5  
6:00 206.5 F 12.3  
8:00 217.7 F 6.7  
10:00 232.4 F 12.2  
12:00 234.7 F 9.4  
14:00 236.5 F 8.5  
16:00 248.3 F 14.5  
18:00 261.0 F 21.5  
20:00 272.6 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.3  
1:00 272.6 F 29.0  
2:00 271.3 F 28.5  
3:00 272.1 F 28.6  
4:00 272.1 F 29.0

Min 271.2 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.4 F 0.0  
5:00 232.2 F 0.0  
10:00 224.5 F 0.0  
15:00 228.0 F 0.0  
20:00 229.4 F 0.0  
25:00 229.8 F 0.0  
30:00 227.4 F 0.0  
Drying Complete

FILLING: 1:23  
HEATING: 20:15  
STERILIZING: 4:00  
VENTING: 2:59  
DRYING: 30:00  
TOTAL CYCLE: 00:58:38  
CYCLE COMPLETE

02/26/2021 12:36 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_