
Midmark M11 - v1.0.0
Total Cycles: 1352

Name: RITTER 1
Sterilizer ID: V2178198
11/10/2021 11:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.5 F 0.0
2:00 189.1 F 3.2
4:00 187.3 F 7.0
6:00 210.8 F 11.6
8:00 221.6 F 5.9
10:00 234.7 F 10.6
12:00 234.9 F 7.4
14:00 244.0 F 12.8
16:00 241.5 F 11.0
18:00 253.8 F 17.2
20:00 264.7 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.7 F 28.9
2:00 271.7 F 28.2
3:00 271.9 F 28.7
4:00 272.5 F 28.8

Min 271.3 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.9 F 0.0
5:00 223.9 F 0.0
10:00 245.3 F 0.0
15:00 238.3 F 0.0
20:00 227.0 F 0.0
25:00 219.9 F 0.0
30:00 218.3 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 21:60
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:47
CYCLE COMPLETE

11/10/2021 12:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____