
Midmark M11 - v1.0.5
Total Cycles: 428

Name: RITTER2
Sterilizer ID: V2313228
10/07/2021 02:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.1 F 0.0
2:00 171.5 F 3.5
4:00 186.4 F 8.4
6:00 209.7 F 9.1
8:00 224.6 F 9.4
10:00 231.5 F 9.3
12:00 241.5 F 13.0
14:00 242.0 F 12.3
16:00 255.3 F 19.1
18:00 266.8 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.4
1:00 272.1 F 29.4
2:00 272.3 F 29.7
3:00 272.4 F 30.1
4:00 271.6 F 29.6

Min 271.5 F 29.2
Max 272.5 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.6 F 0.0
5:00 210.6 F 0.0
10:00 199.4 F 0.0
15:00 207.2 F 0.0
20:00 207.2 F 0.0
25:00 205.6 F 0.0
30:00 203.4 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 19:21
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:21
CYCLE COMPLETE

10/07/2021 03:33 PM
ACCEPTED: X
REJECTED:
Operator: U02