
Midmark M11 - v1.0.5
Total Cycles: 3857

Name: RITTER2
Sterilizer ID: V2313228
08/14/2024 10:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.7 F 0.0
2:00 157.9 F 2.2
4:00 168.0 F 5.5
6:00 200.0 F 9.6
8:00 221.6 F 7.6
10:00 225.2 F 7.9
12:00 233.1 F 12.6
14:00 234.4 F 9.7
16:00 238.1 F 9.7
18:00 246.9 F 14.0
20:00 257.8 F 20.2
22:00 268.1 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.4
1:00 272.0 F 29.3
2:00 272.9 F 29.6
3:00 272.4 F 29.8
4:00 271.8 F 29.8

Min 271.5 F 29.3
Max 273.2 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.7 F 0.0
5:00 238.2 F 0.0
10:00 187.1 F 0.0
15:00 184.0 F 0.0
20:00 181.6 F 0.0
25:00 180.0 F 0.0
30:00 178.7 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 23:59
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:02:58
CYCLE COMPLETE

08/14/2024 11:54 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____