
Midmark M11 - v1.0.0
Total Cycles: 4146

Name: RITTER 1
Sterilizer ID: V2178198
12/21/2023 10:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.6 F 0.0
2:00 159.7 F 1.7
4:00 159.2 F 4.8
6:00 176.0 F 8.4
8:00 197.4 F 12.7
10:00 214.6 F 5.1
12:00 225.8 F 9.5
14:00 235.1 F 10.2
16:00 238.6 F 10.9
18:00 235.9 F 8.5
20:00 249.3 F 14.8
22:00 260.2 F 20.9
24:00 270.2 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.2
1:00 272.6 F 29.0
2:00 271.8 F 28.4
3:00 271.6 F 28.6
4:00 272.6 F 28.9

Min 271.1 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.0 F 0.0
5:00 226.8 F 0.0
10:00 229.3 F 0.0
15:00 231.7 F 0.0
20:00 228.6 F 0.0
25:00 227.5 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 24:55
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:03:33
CYCLE COMPLETE

12/21/2023 11:20 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____