
Midmark M11 - v1.0.0
Total Cycles: 4737

Name: RITTER 1
Sterilizer ID: V2178198
07/16/2024 04:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.5 F 0.0
2:00 145.3 F 2.1
4:00 155.4 F 5.1
6:00 178.5 F 8.3
8:00 202.4 F 12.0
10:00 211.8 F 4.7
12:00 222.1 F 8.8
14:00 231.8 F 12.9
16:00 233.2 F 8.9
18:00 241.2 F 11.7
20:00 243.2 F 12.3
22:00 254.5 F 18.1
24:00 264.9 F 24.2
26:00 271.5 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.4 F 29.5
2:00 271.9 F 29.1
3:00 271.1 F 28.7
4:00 271.9 F 29.4

Min 271.1 F 28.6
Max 272.5 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.7 F 0.0
5:00 215.0 F 0.0
10:00 223.2 F 0.0
15:00 218.0 F 0.0
20:00 211.2 F 0.0
25:00 205.7 F 0.0
30:00 199.6 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 26:06
STERILIZING: 4:00
VENTING: 4:08
DRYING: 30:00
TOTAL CYCLE: 01:05:16
CYCLE COMPLETE

07/16/2024 05:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____