
Midmark M11 - v1.0.0
Total Cycles: 4956

Name: RITTER 1
Sterilizer ID: V2178198
10/02/2024 03:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.8 F 0.0
2:00 164.9 F 0.9
4:00 178.7 F 4.3
6:00 203.5 F 8.3
8:00 213.2 F 12.5
10:00 220.8 F 6.4
12:00 231.5 F 11.0
14:00 231.4 F 7.1
16:00 243.1 F 13.1
18:00 242.3 F 11.5
20:00 254.1 F 17.4
22:00 264.4 F 23.5
24:00 271.9 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.3 F 29.1
2:00 271.6 F 28.4
3:00 271.7 F 28.9
4:00 272.1 F 28.9

Min 271.1 F 28.3
Max 272.4 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.9 F 0.0
5:00 204.0 F 0.0
10:00 198.8 F 0.0
15:00 199.6 F 0.0
20:00 212.9 F 0.0
25:00 225.3 F 0.0
30:00 225.5 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 24:09
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:03:07
CYCLE COMPLETE

10/02/2024 04:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____