
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 004216
14:18 23/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 0:57
PRESSURIZING 0:59
STERILIZING 1:57
136.5 C 222kPa 1:57
136.3 C 223kPa 2:27
136.3 C 221kPa 2:57
136.2 C 221kPa 3:27
136.3 C 222kPa 3:57
136.3 C 222kPa 4:27
136.4 C 223kPa 4:57
136.4 C 222kPa 5:27
136.4 C 222kPa 5:27
VENTING 5:28
AIR DRYING 6:10
CYCLE COMPLETE 17:12

Digital Signature #
141278DEA183649F
