
Midmark M11 - v1.0.0
Total Cycles: 4821

Name: RITTER 1
Sterilizer ID: V2178198
08/15/2024 09:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 125.4 F 0.0
2:00 123.9 F 0.3
4:00 144.9 F 0.8
6:00 166.8 F 1.3
8:00 187.3 F 4.4
10:00 204.3 F 7.8
12:00 218.5 F 11.7
14:00 221.0 F 5.6
16:00 230.4 F 10.2
18:00 234.2 F 8.5
20:00 241.0 F 12.0
22:00 238.7 F 9.9
24:00 251.6 F 16.1
26:00 262.5 F 22.3
28:00 272.2 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 271.3 F 28.4
2:00 272.1 F 28.9
3:00 273.0 F 29.2
4:00 272.3 F 28.8

Min 271.0 F 28.0
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.4 F 0.0
5:00 207.0 F 0.0
10:00 202.7 F 0.0
15:00 210.0 F 0.0
20:00 230.7 F 0.0
25:00 231.8 F 0.0
30:00 226.4 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 28:27
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:06:56
CYCLE COMPLETE

08/15/2024 11:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____