

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3578

Name: RITTER 1  
Sterilizer ID: V2178198  
07/25/2023 11:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 163.8 F 0.0  
2:00 164.8 F 0.8  
4:00 184.8 F 3.3  
6:00 207.8 F 6.9  
8:00 213.8 F 10.9  
10:00 221.3 F 5.9  
12:00 231.6 F 9.3  
14:00 239.0 F 10.9  
16:00 242.0 F 11.8  
18:00 240.0 F 10.3  
20:00 252.6 F 16.5  
22:00 263.4 F 22.8  
24:00 272.5 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.1  
1:00 272.2 F 28.9  
2:00 272.1 F 28.4  
3:00 271.3 F 28.3  
4:00 272.5 F 28.9

Min 271.2 F 28.0  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.0 F 0.0  
5:00 213.1 F 0.0  
10:00 211.9 F 0.0  
15:00 211.4 F 0.0  
20:00 213.3 F 0.0  
25:00 215.7 F 0.0  
30:00 212.0 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 24:09  
STERILIZING: 4:00  
VENTING: 3:38  
DRYING: 30:00  
TOTAL CYCLE: 01:02:44  
CYCLE COMPLETE

07/25/2023 12:30 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_