

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.7uS / 0.4ppm  
CYCLE NUMBER 000296  
9:13 13/11/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:08  
PRESSURIZING 3:11  
STERILIZING 4:20  
136.0 C 229kPa 4:20  
136.4 C 232kPa 4:50  
136.3 C 232kPa 5:20  
136.4 C 233kPa 5:50  
136.1 C 231kPa 6:20  
136.3 C 232kPa 6:50  
136.3 C 233kPa 7:20  
136.4 C 232kPa 7:50  
136.4 C 232kPa 7:50  
VENTING 7:50  
AIR DRYING 8:42  
CYCLE COMPLETE 23:43  
-----

Digital Signature #  
CA3E27F38F11E1F0  
-----