
Midmark M11 - v1.0.0
Total Cycles: 1411

Name: RITTER 1
Sterilizer ID: V2178198
11/30/2021 04:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.0 F 0.0
2:00 186.7 F 4.0
4:00 202.7 F 9.2
6:00 219.6 F 7.1
8:00 232.0 F 9.9
10:00 235.1 F 8.4
12:00 245.3 F 13.6
14:00 244.7 F 12.6
16:00 257.7 F 19.5
18:00 269.2 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.7 F 29.0
2:00 271.5 F 28.2
3:00 272.8 F 29.2
4:00 272.2 F 28.6

Min 271.5 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.4 F 0.0
5:00 209.6 F 0.0
10:00 217.1 F 0.0
15:00 222.7 F 0.0
20:00 227.5 F 0.0
25:00 227.1 F 0.0
30:00 203.6 F 0.0

Drying Complete

FILLING: 2:54
HEATING: 19:03
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:05
CYCLE COMPLETE

12/01/2021 07:42 AM
ACCEPTED: X
REJECTED:
Operator: U02