
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000273
10:51 01/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 3:17
STERILIZING 4:47
136.4 C 233kPa 4:47
136.2 C 232kPa 5:17
136.4 C 233kPa 5:47
136.2 C 233kPa 6:17
136.4 C 233kPa 6:47
136.3 C 233kPa 7:17
136.4 C 233kPa 7:47
136.2 C 232kPa 8:17
136.2 C 232kPa 8:17
VENTING 8:17
AIR DRYING 9:18
CYCLE COMPLETE 24:18

Digital Signature #
5309F5313A1C02A3
