
Midmark M11 - v1.0.5
Total Cycles: 2775

Name: RITTER2
Sterilizer ID: V2313228
08/16/2023 08:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.3 F 0.0
2:00 84.7 F 0.0
4:00 82.2 F 0.0
6:00 100.9 F 0.0
8:00 149.9 F 0.0
10:00 183.3 F 0.0
12:00 206.6 F 0.0
14:00 210.6 F 0.1
16:00 220.8 F 3.3
18:00 228.4 F 6.0
20:00 236.6 F 9.4
22:00 245.3 F 13.4
24:00 240.7 F 10.9
26:00 236.8 F 9.1
28:00 240.2 F 10.1
30:00 248.7 F 14.7
32:00 260.2 F 21.1
34:00 270.6 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.5
1:00 272.5 F 28.8
2:00 272.6 F 29.3
3:00 271.8 F 28.4
4:00 272.7 F 29.1

Min 271.7 F 28.4
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.8 F 0.0
5:00 227.6 F 0.0
10:00 216.2 F 0.0
15:00 197.1 F 0.0
20:00 195.0 F 0.0
25:00 219.5 F 0.0
30:00 227.6 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 34:47
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:12:52
CYCLE COMPLETE

08/16/2023 09:33 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____