
Midmark M11 - v1.0.5
Total Cycles: 2187

Name: RITTER2
Sterilizer ID: V2313228
03/28/2023 01:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 207.2 F 0.0
2:00 198.4 F 4.2
4:00 203.6 F 9.8
6:00 227.9 F 6.7
8:00 236.9 F 11.8
10:00 237.4 F 10.0
12:00 236.3 F 9.0
14:00 251.7 F 16.5
16:00 264.2 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.9
1:00 272.6 F 29.7
2:00 271.9 F 28.9
3:00 272.4 F 29.7
4:00 272.3 F 29.2

Min 271.5 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.4 F 0.0
5:00 233.7 F 0.0
10:00 237.3 F 0.0
15:00 227.1 F 0.0
20:00 229.4 F 0.0
25:00 226.2 F 0.0
30:00 226.7 F 0.0
Drying Complete

FILLING: 2:10
HEATING: 17:46
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:57:12
CYCLE COMPLETE

03/28/2023 02:55 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____