

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1648

Name: RITTER 1  
Sterilizer ID: V2178198  
02/17/2022 11:40 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.4 F 0.0  
2:00 178.3 F 0.0  
4:00 204.4 F 2.3  
6:00 212.3 F 7.5  
8:00 227.8 F 12.8  
10:00 233.4 F 9.3  
12:00 236.6 F 9.2  
14:00 245.0 F 12.7  
16:00 245.5 F 13.0  
18:00 258.8 F 20.1  
20:00 270.4 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.6 F 28.9  
2:00 271.5 F 28.4  
3:00 272.7 F 29.0  
4:00 271.6 F 28.3

Min 271.4 F 28.2  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.1 F 0.0  
5:00 229.5 F 0.0  
10:00 227.2 F 0.0  
15:00 228.1 F 0.0  
20:00 230.7 F 0.0  
25:00 230.7 F 0.0  
30:00 228.7 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 20:48  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:58:58  
CYCLE COMPLETE

02/17/2022 12:42 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02