
Midmark M11 - v1.0.5
Total Cycles: 1085

Name: RITTER2
Sterilizer ID: V2313228
05/09/2022 01:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.6 F 0.0
2:00 160.3 F 2.9
4:00 177.0 F 7.1
6:00 200.9 F 12.3
8:00 220.7 F 7.2
10:00 233.8 F 12.5
12:00 236.4 F 10.0
14:00 235.8 F 8.6
16:00 249.8 F 15.4
18:00 261.6 F 22.3
20:00 272.6 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.2
1:00 272.2 F 29.7
2:00 273.0 F 29.7
3:00 271.7 F 29.1
4:00 272.1 F 29.7

Min 271.4 F 29.0
Max 273.2 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.3 F 0.0
5:00 249.6 F 0.0
10:00 195.5 F 0.0
15:00 191.2 F 0.0
20:00 188.5 F 0.0
25:00 188.4 F 0.0
30:00 193.5 F 0.0
Drying Complete

FILLING: 1:38
HEATING: 21:05
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:54
CYCLE COMPLETE

05/09/2022 02:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____