

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4143

Name: RITTER2  
Sterilizer ID: V2313228  
11/20/2024 03:54 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 148.8 F 0.0  
2:00 144.3 F 0.1  
4:00 158.9 F 2.0  
6:00 196.6 F 5.4  
8:00 209.0 F 9.6  
10:00 230.7 F 9.6  
12:00 229.5 F 8.9  
14:00 240.1 F 12.3  
16:00 240.5 F 11.6  
18:00 238.9 F 10.3  
20:00 251.6 F 16.5  
22:00 262.3 F 22.7  
24:00 272.1 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.4  
1:00 271.8 F 29.8  
2:00 272.7 F 30.2  
3:00 272.2 F 29.7  
4:00 271.4 F 29.3

Min 271.4 F 29.3  
Max 272.9 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.6 F 0.0  
5:00 226.4 F 0.0  
10:00 230.5 F 0.0  
15:00 229.5 F 0.0  
20:00 227.2 F 0.0  
25:00 230.3 F 0.0  
30:00 227.5 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 25:15  
STERILIZING: 4:00  
VENTING: 3:48  
DRYING: 30:00  
TOTAL CYCLE: 01:04:04  
CYCLE COMPLETE

11/20/2024 04:58 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_