

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001823  
7:36 11/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:33  
PRESSURIZING 2:39  
STERILIZING 3:44  
136.0 C 220kPa 3:44  
135.8 C 219kPa 4:14  
136.3 C 221kPa 4:44  
136.2 C 223kPa 5:14  
136.4 C 225kPa 5:44  
136.3 C 223kPa 6:14  
136.3 C 220kPa 6:44  
136.3 C 223kPa 7:14  
136.3 C 223kPa 7:14  
VENTING 7:14  
AIR DRYING 8:04  
CYCLE COMPLETE 9:21  
-----

Digital Signature #  
BC6CD78B8AFF489A  
-----