
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000749
8:30 27/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:18
PRESSURIZING 3:20
STERILIZING 4:30
136.5 C 232kPa 4:30
136.3 C 232kPa 5:00
136.3 C 232kPa 5:30
136.4 C 232kPa 6:00
136.4 C 233kPa 6:30
136.3 C 232kPa 7:00
136.2 C 232kPa 7:30
136.3 C 232kPa 8:00
136.3 C 232kPa 8:00
VENTING 8:01
AIR DRYING 8:55
CYCLE COMPLETE 8:59

Digital Signature #
0618B32C1D0410EB
